Cryostat Users

After using this instrument it must be:
Switched OFF with door OPEN
or
Set to 6°C with the door CLOSED

Under NO CIRCUMSTANCES should the instrument be left:

- with a disposable blade in the blade holder, regardless of temperature
- with samples on the stage or chuck
- at freezing temperatures
- off with the door closed
- at 6°C with the door open

User Name	Date	Time to Freezing Temp	Time to 6°C or OFF	Removed samples and cleaned up excess OCT?